

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001168  
11:49 13/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:33  
PRESSURIZING 3:36  
STERILIZING 5:18  
136.5 C 225kPa 5:18  
136.2 C 225kPa 5:48  
136.3 C 225kPa 6:18  
136.4 C 226kPa 6:48  
136.3 C 224kPa 7:18  
136.3 C 225kPa 7:48  
136.4 C 227kPa 8:18  
136.2 C 224kPa 8:48  
136.2 C 224kPa 8:48  
VENTING 8:48  
AIR DRYING 9:47  
CYCLE COMPLETE 24:47  
-----

Digital Signature #  
5CA6AA730A3FA2DA  
-----