
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007057
6:59 11/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:32
PRESSURIZING 1:35
STERILIZING 2:36
136.0 C 220kPa 2:36
135.6 C 219kPa 3:06
136.4 C 220kPa 3:36
136.1 C 221kPa 4:06
136.4 C 220kPa 4:36
136.3 C 221kPa 5:06
136.1 C 220kPa 5:36
136.2 C 221kPa 6:06
136.2 C 221kPa 6:06
VENTING 6:07
AIR DRYING 6:58
CYCLE COMPLETE 7:04

Digital Signature #
28F14484D3B0D4C4
