
Midmark M11 - v1.0.0
Total Cycles: 5384

Name: RITTER 1
Sterilizer ID: V2178198
03/04/2025 11:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.2 F 0.0
2:00 137.5 F 1.4
4:00 142.6 F 4.5
6:00 166.3 F 8.0
8:00 189.7 F 11.9
10:00 209.2 F 3.3
12:00 214.2 F 7.1
14:00 226.2 F 11.1
16:00 230.0 F 6.6
18:00 241.1 F 11.8
20:00 238.0 F 9.5
22:00 250.4 F 15.5
24:00 261.2 F 21.7
26:00 271.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.7 F 29.4
2:00 271.6 F 28.9
3:00 272.6 F 29.5
4:00 271.7 F 28.8

Min 271.6 F 28.8
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 197.4 F 0.0
10:00 204.0 F 0.0
15:00 205.0 F 0.0
20:00 202.9 F 0.0
25:00 201.7 F 0.0
30:00 203.0 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 27:26
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:06:06
CYCLE COMPLETE

03/04/2025 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____