
Midmark M11 - v1.0.5
Total Cycles: 221

Name: RITTER2
Sterilizer ID: V2313228
07/19/2021 09:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.0 F 0.0
2:00 180.7 F 1.5
4:00 191.0 F 5.8
6:00 211.2 F 10.6
8:00 219.0 F 5.8
10:00 232.4 F 11.1
12:00 232.8 F 8.7
14:00 240.0 F 10.8
16:00 246.9 F 14.1
18:00 258.7 F 20.6
20:00 269.2 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.0
1:00 272.6 F 29.8
2:00 271.6 F 28.9
3:00 272.6 F 29.6
4:00 271.6 F 28.9

Min 271.5 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.8 F 0.0
5:00 214.7 F 0.0
10:00 226.2 F 0.0
15:00 202.8 F 0.0
20:00 188.0 F 0.0
25:00 172.3 F 0.0
30:00 178.3 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:25
CYCLE COMPLETE

07/19/2021 10:16 AM
ACCEPTED: X
REJECTED:
Operator: U01