

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4591

Name: RITTER 1  
Sterilizer ID: V2178198  
05/28/2024 03:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.6 F 0.0  
2:00 166.8 F 3.3  
4:00 179.7 F 7.7  
6:00 218.5 F 13.0  
8:00 226.3 F 7.5  
10:00 233.7 F 13.0  
12:00 236.1 F 9.8  
14:00 235.6 F 8.4  
16:00 248.2 F 14.4  
18:00 260.4 F 21.1  
20:00 270.8 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 271.6 F 28.9  
2:00 271.9 F 29.1  
3:00 272.0 F 29.1  
4:00 272.0 F 29.2

Min 271.3 F 28.5  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.4 F 0.0  
5:00 232.9 F 0.0  
10:00 208.7 F 0.0  
15:00 195.2 F 0.0  
20:00 191.2 F 0.0  
25:00 195.3 F 0.0  
30:00 208.4 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 20:30  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

05/28/2024 04:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_