
Midmark M11 - v1.0.5
Total Cycles: 1116

Name: RITTER2
Sterilizer ID: V2313228
05/25/2022 01:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.5 F 0.0
2:00 180.6 F 3.6
4:00 194.1 F 8.4
6:00 211.2 F 9.6
8:00 229.8 F 9.7
10:00 232.3 F 8.5
12:00 243.0 F 13.5
14:00 243.7 F 12.6
16:00 256.9 F 19.6
18:00 268.9 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.0
1:00 271.8 F 29.0
2:00 273.1 F 30.0
3:00 271.8 F 29.7
4:00 272.1 F 29.5

Min 271.0 F 28.9
Max 273.4 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 231.5 F 0.0
10:00 225.0 F 0.0
15:00 231.2 F 0.0
20:00 225.7 F 0.0
25:00 227.4 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

05/25/2022 02:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____