
Midmark M11 - v1.0.5
Total Cycles: 2679

Name: RITTER2
Sterilizer ID: V2313228
07/21/2023 10:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.9 F 0.0
2:00 145.2 F 0.0
4:00 146.4 F 0.1
6:00 171.9 F 2.9
8:00 197.3 F 6.3
10:00 219.8 F 10.3
12:00 224.4 F 7.3
14:00 231.4 F 9.9
16:00 236.3 F 9.9
18:00 240.1 F 12.1
20:00 239.5 F 10.7
22:00 251.3 F 16.7
24:00 261.8 F 22.7
26:00 271.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.6 F 30.0
2:00 272.0 F 29.4
3:00 271.2 F 29.2
4:00 272.2 F 30.0

Min 271.1 F 29.1
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 211.5 F 0.0
10:00 213.4 F 0.0
15:00 216.5 F 0.0
20:00 220.3 F 0.0
25:00 224.0 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 26:44
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:05:04
CYCLE COMPLETE

07/21/2023 11:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____