
Midmark M11 - v1.0.5
Total Cycles: 337

Name: RITTER2
Sterilizer ID: V2313228
09/08/2021 08:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.4 F 0.0
2:00 125.2 F 1.0
4:00 126.7 F 3.0
6:00 148.3 F 6.4
8:00 175.9 F 10.7
10:00 210.6 F 4.8
12:00 218.9 F 9.2
14:00 228.3 F 10.9
16:00 238.9 F 11.5
18:00 238.1 F 9.9
20:00 250.6 F 15.9
22:00 260.9 F 22.0
24:00 270.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.2
1:00 272.3 F 30.1
2:00 272.5 F 30.3
3:00 271.9 F 30.5
4:00 272.1 F 30.2

Min 271.4 F 30.1
Max 273.3 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.9 F 0.0
5:00 233.6 F 0.0
10:00 225.2 F 0.0
15:00 230.9 F 0.0
20:00 223.8 F 0.0
25:00 226.0 F 0.0
30:00 175.1 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:48
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:04:01
CYCLE COMPLETE

09/08/2021 09:24 AM
ACCEPTED: X
REJECTED:
Operator: U01