
Midmark M11 - v1.0.0
Total Cycles: 2744

Name: RITTER 1
Sterilizer ID: V2178198
01/05/2023 03:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.3 F 0.0
2:00 167.2 F 0.0
4:00 188.5 F 1.6
6:00 207.7 F 6.0
8:00 222.5 F 10.7
10:00 226.4 F 6.2
12:00 237.5 F 11.6
14:00 235.8 F 9.1
16:00 239.1 F 9.6
18:00 247.8 F 14.2
20:00 259.8 F 20.9
22:00 270.7 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 28.9
2:00 271.7 F 28.5
3:00 272.6 F 28.9
4:00 271.6 F 28.4

Min 271.5 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.7 F 0.0
5:00 207.9 F 0.0
10:00 223.9 F 0.0
15:00 228.3 F 0.0
20:00 227.8 F 0.0
25:00 227.1 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

01/05/2023 04:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____