
Midmark M11 - v1.0.0
Total Cycles: 2153

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2022 07:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 76.3 F 0.0
4:00 83.2 F 0.0
6:00 106.0 F 0.0
8:00 134.6 F 1.3
10:00 160.7 F 4.4
12:00 180.5 F 7.8
14:00 197.5 F 11.2
16:00 214.9 F 4.7
18:00 222.3 F 7.5
20:00 232.5 F 11.4
22:00 230.4 F 7.2
24:00 243.2 F 12.8
26:00 240.3 F 10.7
28:00 252.6 F 16.8
30:00 262.9 F 22.9
32:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.7 F 29.4
2:00 272.0 F 28.7
3:00 271.8 F 28.9
4:00 272.6 F 29.3

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 205.8 F 0.0
10:00 211.1 F 0.0
15:00 217.8 F 0.0
20:00 221.6 F 0.0
25:00 228.8 F 0.0
30:00 229.4 F 0.0

Drying Complete

FILLING: 0:45
HEATING: 32:26
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:10:22
CYCLE COMPLETE

07/20/2022 08:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____