
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007768
8:42 20/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:17
STERILIZING 4:08
136.0 C 220kPa 4:08
136.1 C 219kPa 4:38
136.2 C 221kPa 5:08
136.4 C 224kPa 5:38
136.3 C 222kPa 6:08
136.4 C 223kPa 6:38
136.4 C 220kPa 7:08
136.1 C 224kPa 7:38
136.1 C 224kPa 7:38
VENTING 7:38
AIR DRYING 8:30
CYCLE COMPLETE 19:40

Digital Signature #
6BA067D63F073ED2
