
Midmark M11 - v1.0.5
Total Cycles: 580

Name: RITTER2
Sterilizer ID: V2313228
11/29/2021 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.9 F 0.0
2:00 102.8 F 0.0
4:00 114.5 F 0.2
6:00 142.4 F 3.2
8:00 172.5 F 7.2
10:00 192.9 F 12.0
12:00 210.6 F 5.6
14:00 225.5 F 10.1
16:00 231.5 F 8.5
18:00 240.8 F 12.9
20:00 240.9 F 11.8
22:00 253.9 F 18.4
24:00 265.3 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.6
1:00 273.0 F 30.2
2:00 271.8 F 29.4
3:00 272.8 F 30.4
4:00 272.4 F 29.8

Min 271.7 F 29.4
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.8 F 0.0
5:00 207.0 F 0.0
10:00 205.9 F 0.0
15:00 208.8 F 0.0
20:00 208.3 F 0.0
25:00 212.5 F 0.0
30:00 212.5 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 25:47
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:04:06
CYCLE COMPLETE

11/29/2021 11:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____