
Midmark M11 - v1.0.0
Total Cycles: 1623

Name: RITTER 1
Sterilizer ID: V2178198
02/11/2022 11:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.9 F 0.0
2:00 78.8 F 0.0

C103: HEATUP MODE

** ITEMS NOT STERILE **

02/11/2022 11:20 AM