
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004088
10:33 25/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:37
PRESSURIZING 1:42
STERILIZING 2:40
136.4 C 221kPa 2:40
136.4 C 226kPa 3:11
136.4 C 222kPa 3:41
136.1 C 222kPa 4:11
136.3 C 224kPa 4:41
136.4 C 225kPa 5:11
136.4 C 224kPa 5:41
136.4 C 225kPa 6:11
136.4 C 225kPa 6:11
VENTING 6:11
AIR DRYING 7:01
CYCLE COMPLETE 7:41

Digital Signature #
9CB2C951FA4A08F9
