

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3657

Name: RITTER 1  
Sterilizer ID: V2178198  
08/15/2023 02:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
2:00 191.0 F 4.3  
4:00 192.6 F 8.6  
6:00 216.5 F 11.6  
8:00 224.6 F 8.0  
10:00 236.6 F 13.0  
12:00 237.4 F 10.1  
14:00 237.2 F 8.4  
16:00 249.1 F 14.9  
18:00 261.4 F 21.8  
20:00 272.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 271.6 F 28.5  
2:00 271.6 F 28.6  
3:00 271.6 F 28.5  
4:00 271.7 F 28.4  
  
Min 271.6 F 28.4  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 230.7 F 0.0  
10:00 237.3 F 0.0  
15:00 230.4 F 0.0  
20:00 225.0 F 0.0  
25:00 223.0 F 0.0  
30:00 221.3 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 20:17  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:59:12  
CYCLE COMPLETE

08/15/2023 03:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_