

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002655  
10:47 07/04/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:26  
PRESSURIZING 2:32  
STERILIZING 3:38  
136.0 C 220kPa 3:38  
136.4 C 224kPa 4:08  
136.3 C 225kPa 4:38  
136.4 C 225kPa 5:08  
136.1 C 223kPa 5:38  
136.3 C 224kPa 6:08  
136.1 C 223kPa 6:38  
136.4 C 224kPa 7:08  
136.4 C 224kPa 7:08  
VENTING 7:08  
AIR DRYING 8:02  
CYCLE COMPLETE 8:28

-----  
Digital Signature #  
23A7BC9C6E4F25BF  
-----