
Midmark M11 - v1.0.5
Total Cycles: 462

Name: RITTER2
Sterilizer ID: V2313228
10/20/2021 06:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.8 F 0.0
2:00 74.9 F 0.0
4:00 105.9 F 0.0
6:00 124.5 F 0.1
8:00 164.4 F 0.3
10:00 189.7 F 0.7
12:00 203.2 F 4.5
14:00 213.9 F 7.9
16:00 225.0 F 11.8
18:00 228.2 F 7.4
20:00 239.4 F 12.3
22:00 238.2 F 10.2
24:00 236.7 F 8.8
26:00 250.3 F 15.7
28:00 262.0 F 22.4
30:00 272.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.4 F 29.0
2:00 272.0 F 29.3
3:00 271.9 F 28.7
4:00 272.5 F 29.5

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.4 F 0.0
5:00 197.2 F 0.0
10:00 187.3 F 0.0
15:00 184.4 F 0.0
20:00 180.5 F 0.0
25:00 178.5 F 0.0
30:00 173.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:26
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:08:17
CYCLE COMPLETE

10/20/2021 08:26 AM
ACCEPTED: X
REJECTED:
Operator: U01