

-----  
Midmark M11 - v1.0.0  
Total Cycles: 735

Name: RITTER 1  
Sterilizer ID: V2178198  
03/22/2021 09:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 99.3 F 0.0  
2:00 106.7 F 0.0  
4:00 116.4 F 0.2  
6:00 151.7 F 2.6  
8:00 179.9 F 6.5  
10:00 232.1 F 10.8  
12:00 217.6 F 5.5  
14:00 238.2 F 11.1  
16:00 233.1 F 7.4  
18:00 244.7 F 12.7  
20:00 243.4 F 11.6  
22:00 255.8 F 18.1  
24:00 266.5 F 25.1  
26:00 271.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.2 F 29.2  
2:00 272.2 F 28.9  
3:00 271.4 F 28.6  
4:00 271.5 F 28.9

Min 271.2 F 28.6  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.0 F 0.0  
5:00 228.5 F 0.0  
10:00 230.4 F 0.0  
15:00 225.1 F 0.0  
20:00 181.9 F 0.0  
25:00 199.5 F 0.0  
Dry Cycle Aborted

FILLING: 0:51  
HEATING: 26:03  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 27:21  
TOTAL CYCLE: 01:01:10

03/22/2021 10:13 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01