
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002738
10:21 29/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:52
STERILIZING 2:57
136.4 C 220kPa 2:57
136.4 C 224kPa 3:27
136.2 C 219kPa 3:57
136.4 C 224kPa 4:27
136.4 C 224kPa 4:57
136.3 C 221kPa 5:27
136.1 C 221kPa 5:57
136.4 C 221kPa 6:27
136.4 C 221kPa 6:27
VENTING 6:27
AIR DRYING 7:20
CYCLE COMPLETE 7:26

Digital Signature #
17774E3CF70A064E
