
Midmark M11 - v1.0.5
Total Cycles: 4084

Name: RITTER2
Sterilizer ID: V2313228
11/05/2024 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 70.6 F 0.0
4:00 77.6 F 0.0
6:00 102.7 F 0.0
8:00 142.2 F 0.1
10:00 172.1 F 0.3
12:00 191.6 F 1.4
14:00 205.3 F 4.2
16:00 213.0 F 7.2
18:00 219.8 F 10.6
20:00 228.3 F 8.5
22:00 235.0 F 10.3
24:00 236.7 F 9.2
26:00 244.6 F 13.2
28:00 243.0 F 12.0
30:00 254.5 F 17.9
32:00 264.7 F 24.1
34:00 271.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.7 F 29.5
2:00 271.9 F 28.7
3:00 272.1 F 29.3
4:00 272.5 F 29.2

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 197.5 F 0.0
10:00 196.2 F 0.0
15:00 182.2 F 0.0
20:00 181.6 F 0.0
25:00 182.9 F 0.0
30:00 183.8 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 34:06
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:12:28
CYCLE COMPLETE

11/05/2024 11:06 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____