
Midmark M11 - v1.0.0
Total Cycles: 1571

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2022 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.8 F 0.0
2:00 157.5 F 2.4
4:00 173.6 F 6.0
6:00 196.0 F 10.6
8:00 212.1 F 3.9
10:00 226.0 F 8.7
12:00 235.7 F 11.3
14:00 239.6 F 11.1
16:00 237.7 F 9.3
18:00 251.6 F 16.0
20:00 263.5 F 22.9
22:00 271.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.5 F 28.7
2:00 272.0 F 28.7
3:00 272.3 F 28.6
4:00 271.9 F 28.6

Min 271.5 F 28.1
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.0 F 0.0
5:00 228.4 F 0.0
10:00 213.9 F 0.0
15:00 201.6 F 0.0
20:00 195.4 F 0.0
25:00 193.5 F 0.0
30:00 191.4 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 22:03
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

01/31/2022 11:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____