
Midmark M11 - v1.0.5
Total Cycles: 2008

Name: RITTER2
Sterilizer ID: V2313228
02/06/2023 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.9 F 0.0
2:00 76.7 F 0.0
4:00 101.6 F 0.0
6:00 100.9 F 0.0
8:00 140.2 F 0.0
10:00 177.3 F 1.0
12:00 197.9 F 4.3
14:00 213.2 F 7.7
16:00 224.1 F 11.6
18:00 228.5 F 7.2
20:00 239.9 F 12.4
22:00 238.8 F 10.4
24:00 236.4 F 9.0
26:00 250.8 F 15.8
28:00 262.4 F 22.6
30:00 272.8 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.6 F 29.5
2:00 271.6 F 28.9
3:00 272.3 F 29.0
4:00 272.1 F 29.4

Min 271.5 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 232.6 F 0.0
10:00 235.9 F 0.0
15:00 231.6 F 0.0
20:00 225.6 F 0.0
25:00 229.1 F 0.0
30:00 224.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:10
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:08:04
CYCLE COMPLETE

02/06/2023 09:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____