
Midmark M11 - v1.0.5
Total Cycles: 1128

Name: RITTER2
Sterilizer ID: V2313228
05/27/2022 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.2 F 0.0
2:00 202.7 F 3.3
4:00 203.7 F 8.8
6:00 228.0 F 9.2
8:00 234.3 F 11.1
10:00 235.3 F 9.3
12:00 237.9 F 9.5
14:00 249.4 F 15.6
16:00 261.7 F 22.7
18:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.7 F 29.9
2:00 272.2 F 29.3
3:00 271.2 F 29.2
4:00 272.5 F 30.0

Min 271.2 F 28.9
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 222.2 F 0.0
10:00 230.3 F 0.0
15:00 222.3 F 0.0
20:00 219.9 F 0.0
25:00 215.9 F 0.0
30:00 212.5 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 18:25
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

05/27/2022 10:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____