
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002017
8:27 12/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:10
STERILIZING 3:12
136.0 C 220kPa 3:12
136.1 C 223kPa 3:42
136.4 C 225kPa 4:12
136.2 C 222kPa 4:42
136.4 C 223kPa 5:12
136.4 C 225kPa 5:42
136.2 C 222kPa 6:12
136.3 C 224kPa 6:42
136.3 C 224kPa 6:42
VENTING 6:42
AIR DRYING 7:38
CYCLE COMPLETE 29:29

Digital Signature #
4302C6F572005B95
