
Midmark M11 - v1.0.0
Total Cycles: 3590

Name: RITTER 1
Sterilizer ID: V2178198
07/27/2023 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.2 F 0.0
2:00 168.3 F 0.3
4:00 178.4 F 2.9
6:00 198.9 F 6.6
8:00 212.4 F 11.0
10:00 223.0 F 5.0
12:00 230.5 F 10.2
14:00 233.7 F 8.3
16:00 243.3 F 12.7
18:00 241.6 F 11.2
20:00 254.1 F 17.5
22:00 265.0 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.2 F 28.8
2:00 272.0 F 28.9
3:00 271.8 F 28.5
4:00 272.3 F 29.1

Min 271.3 F 28.3
Max 272.4 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 211.8 F 0.0
10:00 227.9 F 0.0
15:00 231.4 F 0.0
20:00 226.1 F 0.0
25:00 223.9 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:53
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:02:08
CYCLE COMPLETE

07/27/2023 12:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____