

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4212

Name: RITTER 1  
Sterilizer ID: V2178198  
01/17/2024 03:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 84.9 F 0.0  
2:00 85.6 F 0.4  
4:00 90.1 F 1.8  
6:00 115.7 F 3.9  
8:00 150.7 F 6.7  
10:00 190.9 F 10.1  
12:00 210.2 F 10.9  
14:00 214.0 F 5.3  
16:00 223.7 F 8.6  
18:00 231.5 F 12.4  
20:00 231.2 F 8.0  
22:00 240.1 F 12.8  
24:00 239.6 F 10.5  
26:00 250.7 F 16.0  
28:00 260.7 F 21.8  
30:00 270.0 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.0  
1:00 271.8 F 29.4  
2:00 272.7 F 29.8  
3:00 271.9 F 29.2  
4:00 271.3 F 29.1  
  
Min 271.2 F 28.9  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 225.1 F 0.0  
10:00 235.3 F 0.0  
15:00 221.5 F 0.0  
20:00 226.1 F 0.0  
25:00 230.5 F 0.0  
30:00 227.4 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 30:53  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:09:45  
CYCLE COMPLETE

01/17/2024 04:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_