
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000727
13:34 29/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:44
STERILIZING 3:49
136.5 C 222kPa 3:49
136.1 C 222kPa 4:19
136.1 C 223kPa 4:49
136.4 C 225kPa 5:19
136.4 C 225kPa 5:49
136.2 C 222kPa 6:19
136.4 C 224kPa 6:49
136.4 C 225kPa 7:19
136.4 C 225kPa 7:19
VENTING 7:19
AIR DRYING 8:13
CYCLE COMPLETE 8:17

Digital Signature #
2A284A5D75A2A1B4
