

-----  
Midmark M11 - v1.0.5  
Total Cycles: 417

Name: RITTER2  
Sterilizer ID: V2313228  
10/05/2021 09:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.9 F 0.0  
2:00 147.9 F 1.6  
4:00 148.4 F 4.5  
6:00 172.7 F 8.9  
8:00 195.9 F 8.9  
10:00 214.0 F 8.2  
12:00 228.2 F 12.1  
14:00 236.5 F 11.4  
16:00 239.0 F 9.9  
18:00 252.1 F 16.6  
20:00 263.7 F 23.4  
22:00 271.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.7 F 29.5  
1:00 272.4 F 29.9  
2:00 272.1 F 29.4  
3:00 272.1 F 29.9  
4:00 272.5 F 29.6

Min 271.5 F 29.4  
Max 273.2 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.2 F 0.0  
5:00 236.4 F 0.0  
10:00 225.8 F 0.0  
15:00 229.0 F 0.0  
20:00 226.8 F 0.0  
25:00 231.9 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:35  
STERILIZING: 4:00  
VENTING: 2:59  
DRYING: 30:00  
TOTAL CYCLE: 01:00:38  
CYCLE COMPLETE

10/05/2021 10:35 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01