
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002460
8:56 28/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:45
STERILIZING 4:53
136.0 C 220kPa 4:53
136.2 C 223kPa 5:23
135.9 C 220kPa 5:53
136.2 C 222kPa 6:23
136.4 C 225kPa 6:53
136.4 C 223kPa 7:23
136.4 C 225kPa 7:53
136.0 C 222kPa 8:23
136.0 C 222kPa 8:23
VENTING 8:23
AIR DRYING 9:19
CYCLE COMPLETE 39:19

Digital Signature #
C2DA0D70DD729325
