
Midmark M11 - v1.0.5
Total Cycles: 2753

Name: RITTER2
Sterilizer ID: V2313228
08/10/2023 01:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.4 F 0.0
2:00 192.0 F 4.4
4:00 208.7 F 10.3
6:00 221.9 F 6.1
8:00 236.0 F 12.3
10:00 237.9 F 10.8
12:00 238.1 F 10.1
14:00 252.9 F 17.4
16:00 265.4 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.7
2:00 271.5 F 29.0
3:00 272.6 F 29.8
4:00 271.9 F 29.2

Min 271.5 F 28.9
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 213.5 F 0.0
10:00 223.0 F 0.0
15:00 227.1 F 0.0
20:00 226.9 F 0.0
25:00 227.5 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 17:41
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:13
CYCLE COMPLETE

08/10/2023 02:07 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____