

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2736

Name: RITTER 1  
Sterilizer ID: V2178198  
01/04/2023 02:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.9 F 0.0  
2:00 183.1 F 0.0  
4:00 203.1 F 4.4  
6:00 221.6 F 9.3  
8:00 233.6 F 9.5  
10:00 236.7 F 10.8  
12:00 236.5 F 8.5  
14:00 240.1 F 9.9  
16:00 248.7 F 14.6  
18:00 261.4 F 21.7  
20:00 272.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.3 F 28.8  
2:00 271.6 F 28.6  
3:00 272.2 F 28.8  
4:00 271.5 F 28.5  
  
Min 271.4 F 28.2  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 229.8 F 0.0  
10:00 242.9 F 0.0  
15:00 236.0 F 0.0  
20:00 226.3 F 0.0  
25:00 224.6 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 20:21  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:31  
CYCLE COMPLETE

01/04/2023 03:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_