

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3494

Name: RITTER2  
Sterilizer ID: V2313228  
04/03/2024 01:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.1 F 0.0  
2:00 152.1 F 2.3  
4:00 160.6 F 5.5  
6:00 190.8 F 9.8  
8:00 206.8 F 7.7  
10:00 220.1 F 7.1  
12:00 232.4 F 11.8  
14:00 234.0 F 9.1  
16:00 239.1 F 10.4  
18:00 247.2 F 14.4  
20:00 259.4 F 21.2  
22:00 270.4 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.1  
1:00 272.9 F 29.8  
2:00 271.8 F 29.0  
3:00 272.8 F 30.0  
4:00 272.4 F 29.5

Min 271.6 F 29.0  
Max 273.2 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.8 F 0.0  
5:00 212.5 F 0.0  
10:00 222.4 F 0.0  
15:00 224.2 F 0.0  
20:00 223.4 F 0.0  
25:00 224.2 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 22:51  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:01:27  
CYCLE COMPLETE

04/03/2024 02:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_