

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1931

Name: RITTER 1  
Sterilizer ID: V2178198  
05/11/2022 12:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 209.8 F 0.0  
2:00 204.0 F 4.2  
4:00 200.7 F 9.0  
6:00 219.7 F 8.9  
8:00 229.0 F 9.0  
10:00 237.9 F 9.9  
12:00 242.0 F 11.8  
14:00 239.9 F 10.2  
16:00 253.5 F 16.9  
18:00 265.2 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.7 F 28.9  
2:00 271.8 F 28.2  
3:00 272.1 F 28.8  
4:00 272.6 F 28.7  
  
Min 271.4 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.6 F 0.0  
5:00 232.6 F 0.0  
10:00 230.9 F 0.0  
15:00 228.1 F 0.0  
20:00 218.2 F 0.0  
25:00 214.8 F 0.0  
30:00 213.7 F 0.0  
Drying Complete

FILLING: 1:31  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:25  
CYCLE COMPLETE

05/11/2022 01:06 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_