

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3973

Name: RITTER2  
Sterilizer ID: V2313228  
09/25/2024 11:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.1 F 0.0  
2:00 100.5 F 0.0  
4:00 112.3 F 0.1  
6:00 142.5 F 1.0  
8:00 172.9 F 3.9  
10:00 185.9 F 7.3  
12:00 201.5 F 11.0  
14:00 223.0 F 5.8  
16:00 230.2 F 9.4  
18:00 236.5 F 11.1  
20:00 240.1 F 11.6  
22:00 238.6 F 10.2  
24:00 251.1 F 16.4  
26:00 261.8 F 22.7  
28:00 272.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.1  
1:00 271.8 F 29.6  
2:00 272.5 F 29.5  
3:00 271.2 F 29.1  
4:00 272.4 F 29.9

Min 271.2 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
5:00 205.2 F 0.0  
10:00 223.6 F 0.0  
15:00 234.4 F 0.0  
20:00 230.4 F 0.0  
25:00 227.8 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 28:25  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:06:59  
CYCLE COMPLETE

09/25/2024 12:28 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_