

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3879

Name: RITTER2  
Sterilizer ID: V2313228  
08/21/2024 12:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.5 F 0.0  
2:00 143.4 F 2.3  
4:00 156.1 F 5.3  
6:00 184.4 F 9.7  
8:00 208.7 F 5.9  
10:00 221.7 F 8.3  
12:00 234.3 F 13.3  
14:00 236.5 F 10.4  
16:00 235.0 F 8.5  
18:00 249.2 F 15.4  
20:00 261.1 F 22.1  
22:00 272.0 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.5 F 29.7  
2:00 271.4 F 29.1  
3:00 272.7 F 29.8  
4:00 271.7 F 29.0

Min 271.3 F 29.0  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.6 F 0.0  
5:00 208.0 F 0.0  
10:00 201.2 F 0.0  
15:00 199.6 F 0.0  
20:00 196.7 F 0.0  
25:00 196.0 F 0.0  
30:00 195.5 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 22:34  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:00:53  
CYCLE COMPLETE

08/21/2024 01:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_