
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001033
8:29 22/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:39
STERILIZING 4:43
136.5 C 223kPa 4:43
136.1 C 222kPa 5:13
136.3 C 223kPa 5:43
136.3 C 222kPa 6:13
135.9 C 220kPa 6:43
135.9 C 220kPa 7:13
136.2 C 222kPa 7:43
136.3 C 223kPa 8:13
136.3 C 223kPa 8:13
VENTING 8:13
AIR DRYING 9:07
CYCLE COMPLETE 9:08

Digital Signature #
9A7A8F65E104AE40
