
Midmark M11 - v1.0.0
Total Cycles: 1942

Name: RITTER 1
Sterilizer ID: V2178198
05/18/2022 07:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.4 F 0.0
2:00 198.4 F 3.0
4:00 207.8 F 6.8
6:00 226.1 F 10.7
8:00 223.0 F 6.1
10:00 234.1 F 9.3
12:00 239.5 F 10.9
14:00 241.1 F 11.3
16:00 237.9 F 9.3
18:00 251.0 F 15.6
20:00 262.2 F 22.0
22:00 272.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 271.7 F 28.1
2:00 272.2 F 28.4
3:00 272.4 F 28.8
4:00 271.5 F 28.3

Min 271.2 F 27.9
Max 272.5 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 217.5 F 0.0
10:00 182.7 F 0.0
15:00 181.5 F 0.0
20:00 213.1 F 0.0
25:00 233.4 F 0.0
30:00 233.3 F 0.0
Drying Complete

FILLING: 0:39
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

05/18/2022 08:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____