
Midmark M11 - v1.0.0
Total Cycles: 3916

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2023 07:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.7 F 0.0
2:00 71.9 F 0.0
4:00 77.1 F 0.0
6:00 104.1 F 0.1
8:00 143.8 F 1.1
10:00 164.0 F 4.6
12:00 191.9 F 8.2
14:00 207.8 F 11.8
16:00 223.2 F 5.1
18:00 230.3 F 9.0
20:00 240.6 F 13.1
22:00 236.3 F 9.3
24:00 239.1 F 9.9
26:00 245.2 F 12.7
28:00 257.0 F 18.9
30:00 267.5 F 25.4
32:00 271.8 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.8 F 28.7
2:00 272.4 F 29.2
3:00 271.4 F 28.4
4:00 272.8 F 28.8

Min 271.3 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.5 F 0.0
5:00 237.8 F 0.0
10:00 233.6 F 0.0
15:00 233.2 F 0.0
20:00 224.9 F 0.0
25:00 228.2 F 0.0
30:00 217.9 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 32:06
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:10:16
CYCLE COMPLETE

10/18/2023 08:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____