
Midmark M11 - v1.0.5
Total Cycles: 4310

Name: RITTER2
Sterilizer ID: V2313228
01/29/2025 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.7 F 0.0
2:00 157.2 F 1.9
4:00 157.8 F 5.2
6:00 179.5 F 9.7
8:00 199.4 F 7.2
10:00 221.1 F 7.6
12:00 228.8 F 12.4
14:00 232.7 F 9.4
16:00 238.1 F 10.3
18:00 244.4 F 13.7
20:00 256.1 F 19.8
22:00 266.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.8
1:00 272.0 F 30.6
2:00 272.2 F 30.2
3:00 271.4 F 29.7
4:00 271.3 F 30.0

Min 271.1 F 29.6
Max 272.5 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 213.5 F 0.0
10:00 213.9 F 0.0
15:00 212.1 F 0.0
20:00 208.0 F 0.0
25:00 208.2 F 0.0
30:00 208.3 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:02:30
CYCLE COMPLETE

01/29/2025 01:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____