
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003569
10:28 02/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:38
STERILIZING 3:42
136.5 C 220kPa 3:42
136.4 C 225kPa 4:12
136.4 C 224kPa 4:42
136.3 C 223kPa 5:12
136.3 C 223kPa 5:42
136.1 C 221kPa 6:12
136.2 C 221kPa 6:42
136.4 C 224kPa 7:12
136.4 C 224kPa 7:12
VENTING 7:13
AIR DRYING 8:04
CYCLE COMPLETE 8:06

Digital Signature #
EF455B58C1647C7B
