
Midmark M11 - v1.0.0
Total Cycles: 2508

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2022 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.1 F 0.0
2:00 133.6 F 1.4
4:00 174.8 F 3.8
6:00 201.6 F 7.7
8:00 206.4 F 13.0
10:00 219.8 F 5.9
12:00 231.8 F 10.4
14:00 235.5 F 7.6
16:00 240.4 F 11.7
18:00 238.2 F 9.5
20:00 251.9 F 16.2
22:00 263.5 F 23.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.4
1:00 272.2 F 29.1
2:00 272.2 F 28.7
3:00 271.6 F 28.6
4:00 272.7 F 29.2

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 219.4 F 0.0
10:00 224.9 F 0.0
15:00 232.8 F 0.0
20:00 230.0 F 0.0
25:00 227.7 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:58
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

11/03/2022 09:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____