
Midmark M11 - v1.0.0
Total Cycles: 1849

Name: RITTER 1
Sterilizer ID: V2178198
04/20/2022 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.1 F 0.0
2:00 184.4 F 4.3
4:00 204.4 F 9.6
6:00 222.1 F 6.3
8:00 233.4 F 10.1
10:00 233.9 F 7.8
12:00 246.2 F 13.1
14:00 245.3 F 12.5
16:00 258.3 F 19.5
18:00 269.9 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.4 F 28.4
2:00 271.5 F 28.2
3:00 272.2 F 28.3
4:00 271.3 F 28.1

Min 271.2 F 27.8
Max 272.5 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 208.6 F 0.0
10:00 210.9 F 0.0
15:00 215.5 F 0.0
20:00 224.3 F 0.0
25:00 231.5 F 0.0
30:00 231.5 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:33
CYCLE COMPLETE

04/20/2022 01:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____