
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002423
14:37 14/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:44
STERILIZING 3:50
136.5 C 224kPa 3:50
136.0 C 221kPa 4:20
136.4 C 223kPa 4:50
136.4 C 225kPa 5:20
136.4 C 225kPa 5:50
136.4 C 224kPa 6:20
136.2 C 223kPa 6:50
136.4 C 226kPa 7:20
136.4 C 226kPa 7:20
VENTING 7:20
AIR DRYING 8:11
CYCLE COMPLETE 8:40

Digital Signature #
B57263E7870BD210
