
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001723
14:05 09/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:05
STERILIZING 4:02
136.4 C 223kPa 4:02
136.4 C 225kPa 4:32
136.3 C 225kPa 5:02
136.3 C 221kPa 5:32
136.2 C 219kPa 6:02
136.4 C 225kPa 6:32
136.4 C 225kPa 7:02
136.1 C 220kPa 7:32
136.1 C 220kPa 7:32
VENTING 7:33
AIR DRYING 8:22
CYCLE COMPLETE 8:33

Digital Signature #
F4C0A59328F19594
