
Midmark M11 - v1.0.0
Total Cycles: 3801

Name: RITTER 1
Sterilizer ID: V2178198
09/20/2023 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.6 F 0.0
2:00 160.3 F 0.0
4:00 173.5 F 1.8
6:00 195.0 F 5.1
8:00 211.9 F 9.2
10:00 226.1 F 11.8
12:00 231.7 F 8.7
14:00 239.3 F 13.6
16:00 241.8 F 11.2
18:00 238.4 F 9.5
20:00 251.7 F 15.8
22:00 262.5 F 22.1
24:00 272.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.0
1:00 272.4 F 28.5
2:00 271.2 F 28.2
3:00 272.4 F 28.4
4:00 271.8 F 28.6

Min 271.0 F 27.9
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.4 F 0.0
5:00 213.4 F 0.0
10:00 201.4 F 0.0
15:00 211.8 F 0.0
20:00 234.9 F 0.0
25:00 230.8 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 24:22
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:02:54
CYCLE COMPLETE

09/20/2023 02:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____