

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000077  
12:21 12/02/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:25  
PRESSURIZING 3:28  
STERILIZING 4:35  
136.0 C 220kPa 4:35  
136.4 C 222kPa 5:05  
136.2 C 220kPa 5:35  
136.2 C 224kPa 6:05  
135.9 C 222kPa 6:35  
136.3 C 223kPa 7:05  
136.3 C 221kPa 7:35  
136.4 C 225kPa 8:05  
136.4 C 225kPa 8:05  
VENTING 8:05  
AIR DRYING 8:58  
CYCLE COMPLETE 8:59

-----  
Digital Signature #  
12B388F2EE452B59  
-----