
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000804
16:15 17/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 4:28
STERILIZING 6:10
136.4 C 223kPa 6:10
136.4 C 223kPa 6:40
136.3 C 222kPa 7:10
136.4 C 222kPa 7:40
136.3 C 222kPa 8:10
136.4 C 223kPa 8:40
136.2 C 221kPa 9:10
136.4 C 223kPa 9:40
136.4 C 223kPa 9:40
VENTING 9:41
AIR DRYING 10:46
CYCLE COMPLETE 31:28

Digital Signature #
9141FB92D5C6C781
