
Midmark M11 - v1.0.0
Total Cycles: 4544

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2024 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 172.1 F 3.6
4:00 187.6 F 7.5
6:00 208.1 F 11.5
8:00 217.5 F 5.5
10:00 226.2 F 10.1
12:00 232.7 F 8.5
14:00 240.7 F 11.9
16:00 238.9 F 10.2
18:00 251.4 F 16.3
20:00 262.2 F 22.4
22:00 272.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.3
2:00 272.4 F 29.0
3:00 271.4 F 28.5
4:00 271.2 F 28.7

Min 271.1 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 223.4 F 0.0
10:00 234.0 F 0.0
15:00 229.2 F 0.0
20:00 230.2 F 0.0
25:00 229.0 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:01:56
CYCLE COMPLETE

05/10/2024 03:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____