
Midmark M11 - v1.0.0
Total Cycles: 3806

Name: RITTER 1
Sterilizer ID: V2178198
09/25/2023 08:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 86.2 F 0.0
4:00 114.3 F 0.0
6:00 106.2 F 0.0
8:00 141.0 F 1.6
10:00 163.6 F 3.2
12:00 204.3 F 5.7
14:00 204.8 F 9.4
16:00 215.7 F 13.0
18:00 222.6 F 5.8
20:00 233.1 F 9.6
22:00 238.6 F 11.1
24:00 239.5 F 10.6
26:00 236.6 F 8.1
28:00 248.7 F 14.3
30:00 259.7 F 20.3
32:00 269.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.2 F 28.3
2:00 272.2 F 28.7
3:00 271.7 F 28.0
4:00 272.6 F 28.7

Min 271.4 F 27.9
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 212.6 F 0.0
10:00 236.7 F 0.0
15:00 223.8 F 0.0
20:00 217.5 F 0.0
25:00 221.8 F 0.0
30:00 225.7 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:01
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:11:06
CYCLE COMPLETE

09/25/2023 10:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____