
Midmark M11 - v1.0.0
Total Cycles: 3181

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2023 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.0 F 0.0
2:00 190.4 F 3.6
4:00 201.8 F 8.3
6:00 220.2 F 9.5
8:00 230.3 F 9.7
10:00 233.7 F 8.7
12:00 243.8 F 12.7
14:00 244.4 F 13.0
16:00 257.8 F 20.1
18:00 269.9 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.3 F 29.2
2:00 271.9 F 29.2
3:00 272.0 F 29.1
4:00 271.9 F 29.3

Min 271.4 F 28.7
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 225.5 F 0.0
10:00 235.2 F 0.0
15:00 233.2 F 0.0
20:00 225.2 F 0.0
25:00 223.7 F 0.0
30:00 223.8 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 18:53
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

05/01/2023 11:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____